

CHIPPEWA CREE TRIBE TRIBAL WATER RESOURCES DEPARTMENT (TWRD)

HOBO TEMPERATURE DATALOGGER “CALIBRATION”⁽¹⁾ LOG

JOB NO: _____ PROJECT NAME: _____

STATE: MT COUNTY: Hill



DATE / TIME: _____ / _____

PERFORMED BY: _____

INSTRUMENT BRAND: _____

MAKE / MODEL: _____

SERIAL NO: _____

CALIBRATION “STANDARD”
TEMPERATURE: _____

CAL STANDARD LOT NO: _____ (if applicable)

RESPONSE TO CAL. “STANDARD”: _____

INSTRUMENT RESPONSE ACCEPTABLE: YES () NO ()

POST-USE RESPONSE TO CAL STANDARD: _____

REMARKS: _____

NOTES:

(1) The HOBO datalogger is not “calibrate-able” (i.e., resettable). Its calibration is a “systems” check against an ice bath “standard” with an NIST-certified thermometer. If the thermometer is within 0.5 dC of NIST thermometer, it passes the test and is acceptable to use in the field.